
Midmark M11 - v1.0.0
Total Cycles: 1639

Name: RITTER 1
Sterilizer ID: V2178198
02/16/2022 08:21 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 75.4 F 0.0
2:00 76.4 F 0.0
4:00 91.1 F 0.0
6:00 119.6 F 0.8
8:00 151.3 F 2.4
10:00 175.1 F 4.8
12:00 192.1 F 7.8
14:00 206.9 F 11.3
16:00 218.9 F 4.1
18:00 228.5 F 8.4
20:00 238.5 F 12.8
22:00 237.5 F 9.7
24:00 238.2 F 9.0
26:00 248.1 F 14.0
28:00 259.7 F 20.4
30:00 270.2 F 27.2

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.0
1:00 272.5 F 28.5
2:00 271.8 F 28.4
3:00 272.1 F 28.3
4:00 272.2 F 28.6

Min 271.4 F 27.9
Max 272.7 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.9 F 0.0
5:00 207.2 F 0.0
10:00 201.2 F 0.0
15:00 199.8 F 0.0
20:00 197.0 F 0.0
25:00 194.7 F 0.0
30:00 219.1 F 0.0
Drying Complete

FILLING: 0:48
HEATING: 30:51
STERILIZING: 4:00
VENTING: 3:02
DRYING: 30:00
TOTAL CYCLE: 01:08:41
CYCLE COMPLETE

02/16/2022 09:29 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____