
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002117
11:54 03/10/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:12
PRESSURIZING 2:17
STERILIZING 3:19
136.0 C 220kPa 3:19
136.4 C 225kPa 3:49
136.4 C 226kPa 4:19
136.4 C 225kPa 4:49
136.4 C 225kPa 5:19
136.3 C 223kPa 5:49
136.4 C 225kPa 6:19
136.4 C 223kPa 6:49
136.4 C 223kPa 6:49
VENTING 6:49
AIR DRYING 7:41
CYCLE COMPLETE 34:35

Digital Signature #
C2F019B441BD26BE
