
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 005196
10:34 06/03/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:07
PRESSURIZING 2:08
STERILIZING 3:12
136.4 C 221kPa 3:12
136.2 C 222kPa 3:42
136.4 C 222kPa 4:12
136.4 C 224kPa 4:42
136.4 C 224kPa 5:12
136.4 C 222kPa 5:42
136.2 C 222kPa 6:12
136.2 C 222kPa 6:42
136.2 C 222kPa 6:42
VENTING 6:42
AIR DRYING 7:32
CYCLE COMPLETE 12:29

Digital Signature #
75EBA1C05A202459
