
Midmark M11 - v1.0.5
Total Cycles: 1669

Name: RITTER2
Sterilizer ID: V2313228
11/02/2022 11:37 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 182.5 F 0.0
2:00 178.9 F 3.5
4:00 186.2 F 7.9
6:00 211.5 F 13.4
8:00 224.9 F 8.5
10:00 235.5 F 11.2
12:00 239.8 F 11.6
14:00 238.8 F 10.3
16:00 252.4 F 17.0
18:00 263.8 F 23.9
20:00 271.6 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.1
1:00 272.6 F 29.8
2:00 271.6 F 29.0
3:00 272.1 F 29.8
4:00 272.5 F 29.5

Min 271.2 F 28.9
Max 272.8 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.3 F 0.0
5:00 217.3 F 0.0
10:00 231.5 F 0.0
15:00 229.0 F 0.0
20:00 225.5 F 0.0
25:00 228.1 F 0.0
30:00 220.9 F 0.0
Drying Complete

FILLING: 1:33
HEATING: 20:06
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 00:58:46
CYCLE COMPLETE

11/02/2022 12:35 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____