
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 000216
14:15 08/07/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:52
PRESSURIZING 2:56
STERILIZING 4:00
136.4 C 223kPa 4:00
136.3 C 224kPa 4:30
136.4 C 223kPa 5:00
136.0 C 218kPa 5:30
136.0 C 221kPa 6:00
136.1 C 220kPa 6:30
136.1 C 222kPa 7:00
136.4 C 225kPa 7:30
136.4 C 225kPa 7:30
VENTING 7:30
AIR DRYING 8:23
CYCLE COMPLETE 8:44

Digital Signature #
96D7BD4B4C204804
