
Midmark M11 - v1.0.0
Total Cycles: 834

Name: RITTER 1
Sterilizer ID: V2178198
05/26/2021 06:52 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.7 F 0.0
2:00 73.9 F 0.0
4:00 90.4 F 0.0
6:00 119.4 F 0.1
8:00 155.5 F 0.2
10:00 185.1 F 0.3
12:00 202.9 F 3.0
14:00 212.6 F 6.1
16:00 222.8 F 9.5
18:00 233.4 F 13.3
20:00 233.3 F 8.7
22:00 243.1 F 12.4
24:00 242.4 F 11.6
26:00 239.6 F 10.0
28:00 253.9 F 17.1
30:00 265.8 F 24.2

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.1
1:00 272.4 F 28.4
2:00 272.4 F 28.8
3:00 271.6 F 28.0
4:00 272.8 F 28.8

Min 271.4 F 27.9
Max 272.8 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 174.0 F 0.0
5:00 205.6 F 0.0
10:00 208.8 F 0.0
15:00 199.7 F 0.0
20:00 197.4 F 0.0
25:00 194.7 F 0.0
30:00 191.5 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 31:38
STERILIZING: 4:00
VENTING: 2:58
DRYING: 30:00
TOTAL CYCLE: 01:09:27
CYCLE COMPLETE

05/26/2021 08:01 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____