
Midmark M11 - v1.0.0
Total Cycles: 4768

Name: RITTER 1
Sterilizer ID: V2178198
07/26/2024 09:57 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 171.9 F 0.0
2:00 170.7 F 1.9
4:00 168.2 F 5.1
6:00 180.1 F 8.6
8:00 203.9 F 12.4
10:00 212.8 F 4.8
12:00 220.9 F 8.7
14:00 233.1 F 12.9
16:00 234.3 F 9.1
18:00 240.5 F 11.0
20:00 244.3 F 12.5
22:00 255.4 F 18.3
24:00 265.5 F 24.4

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.4
1:00 272.4 F 29.1
2:00 271.9 F 28.7
3:00 271.2 F 28.3
4:00 272.0 F 29.1

Min 271.2 F 28.2
Max 272.5 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.1 F 0.0
5:00 204.8 F 0.0
10:00 208.4 F 0.0
15:00 215.2 F 0.0
20:00 227.1 F 0.0
25:00 236.0 F 0.0
30:00 229.2 F 0.0
Drying Complete

FILLING: 0:52
HEATING: 25:57
STERILIZING: 4:00
VENTING: 3:54
DRYING: 30:00
TOTAL CYCLE: 01:04:43
CYCLE COMPLETE

07/26/2024 11:01 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____