

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1104

Name: RITTER2  
Sterilizer ID: V2313228  
05/16/2022 11:43 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 71.9 F 0.0  
2:00 73.7 F 0.0  
4:00 92.2 F 0.0  
6:00 125.9 F 0.0  
8:00 145.9 F 0.3  
10:00 183.9 F 2.5  
12:00 200.1 F 6.6  
14:00 214.3 F 10.8  
16:00 224.1 F 6.6  
18:00 231.9 F 9.8  
20:00 236.7 F 9.9  
22:00 242.1 F 12.3  
24:00 240.5 F 10.9  
26:00 253.3 F 17.4  
28:00 264.6 F 24.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.8  
1:00 272.0 F 28.8  
2:00 272.6 F 29.3  
3:00 272.8 F 29.6  
4:00 271.8 F 29.1

Min 271.6 F 28.6  
Max 272.9 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 226.6 F 0.0  
Dry Cycle Aborted

FILLING: 1:00  
HEATING: 29:51  
STERILIZING: 4:00  
VENTING: 2:41  
DRYING: 4:51  
TOTAL CYCLE: 00:42:24

05/16/2022 12:25 PM  
ACCEPTED:  
REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_