
Midmark M11 - v1.0.0
Total Cycles: 3353

Name: RITTER 1
Sterilizer ID: V2178198
06/07/2023 02:58 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 194.2 F 0.0
2:00 183.3 F 0.5
4:00 207.2 F 5.5
6:00 225.0 F 10.8
8:00 228.5 F 6.6
10:00 239.5 F 12.3
12:00 238.5 F 10.3
14:00 237.3 F 8.7
16:00 251.2 F 15.9
18:00 263.3 F 22.9
20:00 271.6 F 28.2

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.3
1:00 272.4 F 28.9
2:00 271.4 F 28.3
3:00 272.6 F 29.0
4:00 271.7 F 28.3

Min 271.3 F 28.2
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.6 F 0.0
5:00 221.9 F 0.0
10:00 233.8 F 0.0
15:00 230.1 F 0.0
20:00 224.9 F 0.0
25:00 231.1 F 0.0
30:00 221.0 F 0.0
Drying Complete

FILLING: 1:41
HEATING: 20:06
STERILIZING: 4:00
VENTING: 3:19
DRYING: 30:00
TOTAL CYCLE: 00:59:06
CYCLE COMPLETE

06/07/2023 03:57 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____