

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4877

Name: RITTER 1  
Sterilizer ID: V2178198  
09/09/2024 09:13 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 68.8 F 0.0  
2:00 69.3 F 0.0  
4:00 92.1 F 0.0  
6:00 126.5 F 0.0  
8:00 150.2 F 0.1  
10:00 176.0 F 0.3  
12:00 194.5 F 0.8  
14:00 210.0 F 2.4  
16:00 217.6 F 5.1  
18:00 223.4 F 7.8  
20:00 231.8 F 11.1  
22:00 232.4 F 7.4  
24:00 240.5 F 11.4  
26:00 236.7 F 9.1  
28:00 240.5 F 10.1  
30:00 247.5 F 13.8  
32:00 258.2 F 19.5  
34:00 267.5 F 25.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.1  
1:00 271.9 F 28.4  
2:00 272.5 F 28.9  
3:00 272.7 F 28.8  
4:00 272.4 F 28.5  
  
Min 271.5 F 28.0  
Max 272.7 F 28.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 189.3 F 0.0  
5:00 197.7 F 0.0  
10:00 204.3 F 0.0  
15:00 206.3 F 0.0  
20:00 229.9 F 0.0  
25:00 231.5 F 0.0  
30:00 224.5 F 0.0  
Drying Complete

FILLING: 0:40  
HEATING: 35:42  
STERILIZING: 4:00  
VENTING: 3:50  
DRYING: 30:00  
TOTAL CYCLE: 01:14:13  
CYCLE COMPLETE

09/09/2024 10:27 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_