

-----  
Midmark M11 - v1.0.5  
Total Cycles: 990

Name: RITTER2  
Sterilizer ID: V2313228  
04/14/2022 12:45 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 187.1 F 0.0  
2:00 181.1 F 0.7  
4:00 203.1 F 5.6  
6:00 224.6 F 11.3  
8:00 229.5 F 7.9  
10:00 241.4 F 12.3  
12:00 242.0 F 12.6  
14:00 243.0 F 12.3  
16:00 256.9 F 19.6  
18:00 269.0 F 27.3

STERILIZING  
mm:ss Degrees PSI  
0:00 272.0 F 29.1  
1:00 272.5 F 29.9  
2:00 271.9 F 29.1  
3:00 272.5 F 29.9  
4:00 272.3 F 29.4  
  
Min 271.5 F 28.9  
Max 272.7 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 211.8 F 0.0  
5:00 211.7 F 0.0  
10:00 220.8 F 0.0  
15:00 227.2 F 0.0  
20:00 230.4 F 0.0  
25:00 230.1 F 0.0  
30:00 228.6 F 0.0  
Drying Complete

FILLING: 1:53  
HEATING: 18:56  
STERILIZING: 4:00  
VENTING: 3:13  
DRYING: 30:00  
TOTAL CYCLE: 00:58:01  
CYCLE COMPLETE

04/14/2022 01:42 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_