

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2507

Name: RITTER 1  
Sterilizer ID: V2178198  
11/03/2022 06:56 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 80.6 F 0.0  
2:00 81.2 F 0.0  
4:00 85.2 F 0.1  
6:00 109.2 F 2.2  
8:00 142.9 F 5.4  
10:00 174.3 F 9.2  
12:00 220.9 F 13.1  
14:00 220.1 F 4.9  
16:00 229.5 F 8.4  
18:00 238.6 F 12.3  
20:00 234.5 F 8.0  
22:00 245.4 F 13.3  
24:00 242.2 F 11.1  
26:00 254.1 F 17.1  
28:00 264.7 F 23.5  
30:00 270.6 F 28.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.5  
1:00 271.5 F 28.4  
2:00 272.1 F 28.5  
3:00 272.5 F 28.7  
4:00 272.5 F 28.9  
  
Min 271.2 F 28.4  
Max 272.8 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 176.1 F 0.0  
5:00 234.1 F 0.0  
10:00 226.3 F 0.0  
15:00 234.8 F 0.0  
20:00 229.0 F 0.0  
25:00 224.2 F 0.0  
30:00 225.4 F 0.0  
Drying Complete

FILLING: 0:53  
HEATING: 30:38  
STERILIZING: 4:00  
VENTING: 3:08  
DRYING: 30:00  
TOTAL CYCLE: 01:08:38  
CYCLE COMPLETE

11/03/2022 08:04 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_