
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 007278
13:39 26/04/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:19
PRESSURIZING 2:21
STERILIZING 3:22
136.4 C 221kPa 3:22
136.2 C 223kPa 3:52
136.3 C 223kPa 4:22
136.4 C 222kPa 4:52
136.4 C 220kPa 5:22
136.4 C 223kPa 5:52
136.4 C 219kPa 6:22
136.4 C 220kPa 6:52
136.4 C 220kPa 6:52
VENTING 6:52
AIR DRYING 7:44
CYCLE COMPLETE 35:29

Digital Signature #
1CDC987AA62110E1
