
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 007413
6:37 24/05/2024

UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:03
PRESSURIZING 3:05
STERILIZING 4:08
136.4 C 221kPa 4:08
136.4 C 223kPa 4:39
136.3 C 219kPa 5:09
136.4 C 222kPa 5:39
136.2 C 217kPa 6:09
136.3 C 220kPa 6:39
136.4 C 223kPa 7:09
136.4 C 222kPa 7:39
136.4 C 222kPa 7:39
VENTING 7:39
AIR DRYING 8:28
CYCLE COMPLETE 8:34

Digital Signature #
732D492CEB95886B
