
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001462
10:30 20/09/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:49
PRESSURIZING 2:56
STERILIZING 4:09
136.3 C 222kPa 4:09
136.3 C 225kPa 4:39
136.3 C 220kPa 5:09
136.4 C 224kPa 5:39
136.3 C 223kPa 6:09
136.2 C 223kPa 6:39
136.4 C 225kPa 7:09
136.3 C 221kPa 7:39
136.3 C 221kPa 7:39
VENTING 7:39
AIR DRYING 8:28
CYCLE COMPLETE 8:36

Digital Signature #
4BF955A76D43FCB2
