
Surrey, BC
Canada

STATIM 5000 S501R711
SN 511018F00076
Unit # : 000
WATER QUALITY
3.8uS / 2.4ppm
CYCLE NUMBER 000006
8:32 13/01/2020
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CYCLE ABORTED 0:22
STOP BUTTON PRESSED
NOT STERILE
END TIME 0:22

Digital Signature #
3C528048CF7E355D
