
Midmark M11 - v1.0.0
Total Cycles: 1111

Name: RITTER 1
Sterilizer ID: V2178198
08/23/2021 12:54 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 140.3 F 0.0
2:00 142.2 F 1.2
4:00 163.6 F 4.1
6:00 185.6 F 8.1
8:00 208.8 F 13.3
10:00 221.5 F 7.3
12:00 236.9 F 12.2
14:00 235.7 F 8.8
16:00 239.9 F 10.5
18:00 246.1 F 13.1
20:00 258.3 F 19.7
22:00 269.4 F 26.7

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.2
1:00 272.3 F 28.6
2:00 271.7 F 28.5
3:00 272.4 F 28.6
4:00 271.5 F 28.4

Min 271.2 F 28.1
Max 272.6 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 174.5 F 0.0
5:00 216.4 F 0.0
10:00 221.2 F 0.0
15:00 193.6 F 0.0
20:00 190.1 F 0.0
25:00 195.8 F 0.0
30:00 191.0 F 0.0
Drying Complete

FILLING: 1:39
HEATING: 22:59
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 01:01:45
CYCLE COMPLETE

08/23/2021 01:56 PM
ACCEPTED: X
REJECTED:
Operator: U02