
Midmark M11 - v1.0.0
Total Cycles: 2037

Name: RITTER 1
Sterilizer ID: V2178198
06/20/2022 01:47 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 204.5 F 0.0
2:00 205.0 F 3.6
4:00 217.3 F 9.7
6:00 222.2 F 5.9
8:00 236.6 F 11.0
10:00 235.3 F 8.4
12:00 239.9 F 9.8
14:00 248.8 F 14.3
16:00 261.7 F 21.5
18:00 272.7 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 27.9
1:00 272.4 F 28.3
2:00 271.8 F 28.2
3:00 272.2 F 28.2
4:00 271.7 F 28.2

Min 271.3 F 27.7
Max 272.7 F 28.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.5 F 0.0
5:00 215.4 F 0.0
10:00 230.3 F 0.0
15:00 225.6 F 0.0
20:00 220.4 F 0.0
25:00 218.1 F 0.0
30:00 216.3 F 0.0
Drying Complete

FILLING: 2:13
HEATING: 18:18
STERILIZING: 4:00
VENTING: 2:59
DRYING: 30:00
TOTAL CYCLE: 00:57:31
CYCLE COMPLETE

06/20/2022 02:44 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____