
Midmark M11 - v1.0.5
Total Cycles: 1279

Name: RITTER2
Sterilizer ID: V2313228
07/11/2022 08:52 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 191.5 F 0.0
2:00 186.8 F 2.7
4:00 195.8 F 7.0
6:00 211.6 F 12.2
8:00 224.2 F 7.6
10:00 236.6 F 13.1
12:00 236.5 F 10.7
14:00 236.6 F 8.9
16:00 248.8 F 15.5
18:00 260.1 F 21.9
20:00 270.5 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.0 F 29.4
1:00 271.8 F 30.1
2:00 272.8 F 30.3
3:00 272.6 F 30.0
4:00 271.9 F 29.7

Min 271.0 F 29.2
Max 272.8 F 30.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.1 F 0.0
5:00 219.4 F 0.0
10:00 229.3 F 0.0
15:00 246.0 F 0.0
20:00 237.3 F 0.0
25:00 226.7 F 0.0
30:00 226.1 F 0.0
Drying Complete

FILLING: 1:40
HEATING: 20:58
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 00:59:43
CYCLE COMPLETE

07/11/2022 09:51 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____