

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3364

Name: RITTER2  
Sterilizer ID: V2313228  
02/07/2024 12:44 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 158.5 F 0.0  
2:00 157.9 F 2.0  
4:00 174.7 F 5.8  
6:00 194.9 F 10.4  
8:00 213.8 F 5.8  
10:00 222.0 F 9.5  
12:00 232.9 F 9.7  
14:00 241.1 F 12.5  
16:00 241.7 F 11.5  
18:00 254.1 F 18.0  
20:00 265.3 F 24.8  
22:00 271.3 F 29.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 29.4  
1:00 272.7 F 29.6  
2:00 272.1 F 29.9  
3:00 272.0 F 29.4  
4:00 272.3 F 29.9

Min 271.4 F 29.2  
Max 273.0 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 217.4 F 0.0  
5:00 210.3 F 0.0  
10:00 197.7 F 0.0  
15:00 197.2 F 0.0  
20:00 196.4 F 0.0  
25:00 196.3 F 0.0  
30:00 195.9 F 0.0  
Drying Complete

FILLING: 1:11  
HEATING: 22:38  
STERILIZING: 4:00  
VENTING: 3:36  
DRYING: 30:00  
TOTAL CYCLE: 01:01:25  
CYCLE COMPLETE

02/07/2024 01:45 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_