
Midmark M11 - v1.0.5
Total Cycles: 2126

Name: RITTER2
Sterilizer ID: V2313228
03/08/2023 12:14 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 177.2 F 0.0
2:00 174.9 F 3.3
4:00 196.8 F 7.9
6:00 214.2 F 12.3
8:00 227.7 F 8.8
10:00 235.8 F 10.2
12:00 242.9 F 12.9
14:00 243.2 F 12.2
16:00 256.8 F 19.3
18:00 268.7 F 27.0

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.0
1:00 272.2 F 29.1
2:00 272.4 F 29.7
3:00 271.8 F 28.9
4:00 272.4 F 29.6

Min 271.4 F 28.8
Max 272.7 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.8 F 0.0
5:00 225.4 F 0.0
10:00 218.6 F 0.0
15:00 213.1 F 0.0
20:00 216.3 F 0.0
25:00 217.3 F 0.0
30:00 216.5 F 0.0
Drying Complete

FILLING: 1:46
HEATING: 19:04
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:58:00
CYCLE COMPLETE

03/08/2023 01:11 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____