
Midmark M11 - v1.0.5
Total Cycles: 533

Name: RITTER2
Sterilizer ID: V2313228
11/10/2021 10:20 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 95.8 F 0.0
2:00 96.5 F 0.0
4:00 112.5 F 0.1
6:00 144.2 F 1.4
8:00 176.3 F 4.9
10:00 195.4 F 8.8
12:00 209.9 F 12.9
14:00 219.3 F 6.5
16:00 230.5 F 10.8
18:00 231.2 F 7.7
20:00 242.1 F 12.9
22:00 241.0 F 11.4
24:00 252.9 F 17.4
26:00 263.5 F 23.7
28:00 272.6 F 29.5

STERILIZING
mm:ss Degrees PSI
0:00 272.1 F 29.1
1:00 272.2 F 29.8
2:00 272.1 F 29.2
3:00 272.1 F 29.8
4:00 272.4 F 29.4

Min 271.5 F 29.0
Max 273.0 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 180.2 F 0.0
5:00 212.7 F 0.0
10:00 212.7 F 0.0
15:00 216.7 F 0.0
20:00 215.7 F 0.0
25:00 218.0 F 0.0
30:00 221.4 F 0.0
Drying Complete

FILLING: 0:59
HEATING: 28:07
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 01:06:14
CYCLE COMPLETE

11/10/2021 11:26 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____