

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R750  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 000903  
11:40 30/11/2020  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:30  
PRESSURIZING 3:32  
STERILIZING 4:34  
136.4 C 224kPa 4:34  
136.4 C 223kPa 5:04  
136.4 C 222kPa 5:34  
136.4 C 224kPa 6:04  
136.4 C 224kPa 6:34  
136.4 C 222kPa 7:04  
136.4 C 222kPa 7:34  
136.1 C 220kPa 8:04  
136.1 C 220kPa 8:04  
VENTING 8:04  
AIR DRYING 9:00  
CYCLE COMPLETE 9:02

-----  
Digital Signature #  
B90681C35845854A  
-----