
Midmark M11 - v1.0.0
Total Cycles: 3774

Name: RITTER 1
Sterilizer ID: V2178198
09/14/2023 01:38 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 182.9 F 0.0
2:00 180.2 F 2.4
4:00 191.8 F 6.6
6:00 209.1 F 10.7
8:00 221.0 F 5.7
10:00 227.6 F 10.1
12:00 233.2 F 8.3
14:00 242.4 F 12.7
16:00 241.4 F 11.3
18:00 253.8 F 17.5
20:00 264.7 F 24.0
22:00 271.5 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.7
1:00 272.7 F 29.4
2:00 271.9 F 28.8
3:00 271.7 F 29.0
4:00 272.7 F 29.4

Min 271.4 F 28.6
Max 272.8 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.9 F 0.0
5:00 226.2 F 0.0
10:00 226.5 F 0.0
15:00 223.2 F 0.0
20:00 223.0 F 0.0
25:00 222.7 F 0.0
30:00 221.3 F 0.0
Drying Complete

FILLING: 1:34
HEATING: 22:04
STERILIZING: 4:00
VENTING: 3:46
DRYING: 30:00
TOTAL CYCLE: 01:01:24
CYCLE COMPLETE

09/14/2023 02:39 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____