
Midmark M11 - v1.0.0
Total Cycles: 2432

Name: RITTER 1
Sterilizer ID: V2178198
10/14/2022 10:27 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 185.9 F 0.0
2:00 185.1 F 3.9
4:00 189.3 F 7.9
6:00 215.8 F 12.6
8:00 223.2 F 6.2
10:00 235.1 F 11.1
12:00 233.9 F 7.6
14:00 243.8 F 12.8
16:00 243.2 F 12.1
18:00 256.0 F 18.9
20:00 267.2 F 25.8

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.7
1:00 272.7 F 29.4
2:00 271.8 F 28.7
3:00 272.2 F 29.4
4:00 272.4 F 29.2

Min 271.5 F 28.6
Max 272.7 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.7 F 0.0
5:00 220.4 F 0.0
10:00 243.0 F 0.0
15:00 239.9 F 0.0
20:00 228.5 F 0.0
25:00 222.3 F 0.0
30:00 217.5 F 0.0
Drying Complete

FILLING: 0:59
HEATING: 21:35
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:59:45
CYCLE COMPLETE

10/14/2022 11:26 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____