

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3736

Name: RITTER2  
Sterilizer ID: V2313228  
06/26/2024 11:38 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 135.6 F 0.0  
2:00 137.7 F 1.7  
4:00 152.6 F 4.6  
6:00 179.5 F 8.5  
8:00 200.6 F 13.0  
10:00 214.8 F 6.6  
12:00 225.3 F 11.6  
14:00 229.9 F 8.3  
16:00 240.7 F 12.0  
18:00 242.7 F 12.7  
20:00 255.1 F 19.0  
22:00 266.1 F 25.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 29.5  
1:00 272.1 F 30.3  
2:00 272.8 F 30.2  
3:00 272.0 F 29.7  
4:00 271.0 F 29.3

Min 271.0 F 29.3  
Max 273.1 F 30.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 188.1 F 0.0  
5:00 209.7 F 0.0  
10:00 209.7 F 0.0  
15:00 194.4 F 0.0  
20:00 184.8 F 0.0  
25:00 179.8 F 0.0  
30:00 176.4 F 0.0  
Drying Complete

FILLING: 1:02  
HEATING: 23:48  
STERILIZING: 4:00  
VENTING: 3:32  
DRYING: 30:00  
TOTAL CYCLE: 01:02:23  
CYCLE COMPLETE

06/26/2024 12:40 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_