
Midmark M11 - v1.0.0
Total Cycles: 4375

Name: RITTER 1
Sterilizer ID: V2178198
03/12/2024 08:46 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 76.1 F 0.0
2:00 76.3 F 0.2
4:00 87.4 F 1.2
6:00 117.8 F 3.5
8:00 148.5 F 6.0
10:00 200.5 F 9.0
12:00 210.2 F 12.9
14:00 215.6 F 4.1
16:00 218.9 F 7.3
18:00 233.3 F 10.9
20:00 230.9 F 7.3
22:00 239.1 F 10.8
24:00 234.0 F 7.6
26:00 246.6 F 13.6
28:00 257.9 F 19.7
30:00 268.1 F 26.1

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.4
1:00 272.4 F 28.7
2:00 272.1 F 29.1
3:00 271.5 F 28.4
4:00 272.6 F 29.0

Min 271.1 F 28.3
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 172.9 F 0.0
5:00 243.7 F 0.0
10:00 213.4 F 0.0
15:00 223.4 F 0.0
20:00 213.3 F 0.0
25:00 212.3 F 0.0
30:00 207.3 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 31:18
STERILIZING: 4:00
VENTING: 3:27
DRYING: 30:00
TOTAL CYCLE: 01:09:36
CYCLE COMPLETE

03/12/2024 09:55 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____