
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000221
7:46 09/07/2020

UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:04
PRESSURIZING 3:07
STERILIZING 4:13
136.5 C 224kPa 4:13
136.3 C 222kPa 4:43
136.2 C 222kPa 5:13
136.4 C 221kPa 5:43
136.2 C 223kPa 6:13
136.4 C 224kPa 6:43
136.2 C 221kPa 7:13
136.1 C 219kPa 7:43
136.1 C 219kPa 7:43
VENTING 7:43
AIR DRYING 8:33
CYCLE COMPLETE 8:47

Digital Signature #
F86C7DDC7361A622
