

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5118

Name: RITTER 1  
Sterilizer ID: V2178198  
11/21/2024 09:54 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 71.3 F 0.0  
2:00 74.7 F 0.0  
4:00 85.1 F 0.0  
6:00 119.6 F 0.0  
8:00 147.0 F 0.2  
10:00 177.9 F 0.5  
12:00 200.3 F 2.6  
14:00 214.0 F 5.8  
16:00 223.0 F 9.2  
18:00 233.2 F 13.0  
20:00 233.0 F 8.6  
22:00 243.9 F 13.6  
24:00 241.6 F 11.3  
26:00 238.0 F 9.5  
28:00 251.9 F 16.2  
30:00 263.6 F 23.0

STERILIZING  
mm:ss Degrees PSI  
0:00 272.0 F 28.2  
1:00 272.1 F 28.7  
2:00 272.4 F 28.6  
3:00 271.6 F 28.2  
4:00 272.9 F 29.0

Min 271.5 F 28.0  
Max 272.9 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 185.7 F 0.0  
5:00 202.9 F 0.0  
10:00 188.8 F 0.0  
15:00 187.6 F 0.0  
20:00 191.1 F 0.0  
25:00 206.8 F 0.0  
30:00 211.4 F 0.0  
Drying Complete

FILLING: 0:41  
HEATING: 31:57  
STERILIZING: 4:00  
VENTING: 3:24  
DRYING: 30:00  
TOTAL CYCLE: 01:10:03  
CYCLE COMPLETE

11/21/2024 11:04 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_