
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.9uS / 0.5ppm
CYCLE NUMBER 000299
12:56 13/11/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:57
PRESSURIZING 3:59
STERILIZING 5:15
136.0 C 229kPa 5:15
136.4 C 233kPa 5:45
136.4 C 233kPa 6:15
136.1 C 229kPa 6:45
136.4 C 232kPa 7:15
136.2 C 232kPa 7:45
136.3 C 232kPa 8:15
136.4 C 231kPa 8:45
136.4 C 231kPa 8:45
VENTING 8:45
AIR DRYING 9:39
CYCLE COMPLETE 9:42

Digital Signature #
A6575E78B9DFCD2B
