
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 006671
12:32 29/11/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:40
PRESSURIZING 2:47
STERILIZING 3:49
136.0 C 220kPa 3:49
136.4 C 225kPa 4:19
136.4 C 224kPa 4:49
136.1 C 221kPa 5:19
136.4 C 223kPa 5:49
136.4 C 223kPa 6:19
136.4 C 223kPa 6:49
136.1 C 221kPa 7:19
136.1 C 221kPa 7:19
VENTING 7:19
AIR DRYING 8:07
CYCLE COMPLETE 8:12

Digital Signature #
7FE824CFE109137C
