

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2871

Name: RITTER 1  
Sterilizer ID: V2178198  
02/07/2023 03:30 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 174.8 F 0.0  
2:00 176.5 F 3.7  
4:00 182.2 F 8.1  
6:00 205.2 F 13.3  
8:00 221.9 F 7.6  
10:00 235.2 F 12.8  
12:00 236.2 F 10.0  
14:00 235.2 F 8.3  
16:00 249.6 F 15.4  
18:00 262.4 F 22.6  
20:00 271.9 F 28.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.7  
1:00 272.1 F 28.9  
2:00 272.5 F 29.4  
3:00 271.5 F 28.6  
4:00 272.7 F 29.4  
  
Min 271.3 F 28.5  
Max 272.9 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 212.6 F 0.0  
5:00 232.9 F 0.0  
10:00 238.8 F 0.0  
15:00 235.1 F 0.0  
20:00 223.0 F 0.0  
25:00 222.7 F 0.0  
30:00 223.3 F 0.0  
Drying Complete

FILLING: 1:20  
HEATING: 20:08  
STERILIZING: 4:00  
VENTING: 3:16  
DRYING: 30:00  
TOTAL CYCLE: 00:58:44  
CYCLE COMPLETE

02/07/2023 04:28 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_