
Midmark M11 - v1.0.5
Total Cycles: 12

Name: RITTER2
Sterilizer ID: V2313228
04/22/2021 02:31 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 146.4 F 0.0
2:00 144.7 F 2.2
4:00 158.5 F 5.0
6:00 183.8 F 8.6
8:00 201.9 F 12.4
10:00 214.1 F 5.9
12:00 222.6 F 10.3
14:00 232.8 F 9.0
16:00 240.1 F 12.9
18:00 242.4 F 12.3
20:00 253.0 F 18.1
22:00 263.7 F 24.2
24:00 273.0 F 30.6

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.8
1:00 272.5 F 30.6
2:00 272.2 F 30.1
3:00 271.4 F 29.9
4:00 272.2 F 30.7

Min 271.4 F 29.7
Max 272.9 F 30.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.2 F 0.0
5:00 219.6 F 0.0
10:00 217.2 F 0.0
15:00 211.1 F 0.0
20:00 207.1 F 0.0
25:00 204.5 F 0.0
30:00 201.5 F 0.0
Drying Complete

FILLING: 1:18
HEATING: 24:24
STERILIZING: 4:00
VENTING: 3:50
DRYING: 30:00
TOTAL CYCLE: 01:03:33
CYCLE COMPLETE

04/22/2021 03:34 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____