
Midmark M11 - v1.0.0
Total Cycles: 1422

Name: RITTER 1
Sterilizer ID: V2178198
12/06/2021 09:13 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.4 F 0.0
2:00 86.1 F 0.0
4:00 78.0 F 0.0
6:00 98.4 F 0.1
8:00 130.9 F 1.2
10:00 163.9 F 3.1
12:00 184.1 F 5.0
14:00 206.7 F 8.1
16:00 221.4 F 11.7
18:00 223.0 F 5.7
20:00 233.6 F 10.2
22:00 236.3 F 9.0
24:00 242.5 F 12.2
26:00 239.4 F 10.4
28:00 252.9 F 16.9
30:00 263.7 F 23.3
32:00 271.6 F 28.4

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.4
1:00 272.1 F 28.7
2:00 272.5 F 29.2
3:00 271.3 F 28.6
4:00 272.1 F 28.7

Min 271.1 F 28.3
Max 272.6 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 192.2 F 0.0
5:00 210.9 F 0.0
10:00 212.6 F 0.0
15:00 206.6 F 0.0
20:00 185.7 F 0.0
25:00 170.7 F 0.0
30:00 165.2 F 0.0

Drying Complete

FILLING: 0:59
HEATING: 32:06
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 01:10:14
CYCLE COMPLETE

12/06/2021 10:40 AM
ACCEPTED: X
REJECTED:
Operator: U02