

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 000587  
14:36 16/08/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:45  
PRESSURIZING 3:48  
STERILIZING 5:07  
136.4 C 233kPa 5:07  
136.2 C 231kPa 5:37  
136.3 C 232kPa 6:07  
136.3 C 232kPa 6:37  
136.3 C 231kPa 7:07  
136.2 C 231kPa 7:37  
136.3 C 231kPa 8:07  
136.3 C 231kPa 8:37  
136.3 C 231kPa 8:37  
VENTING 8:37  
AIR DRYING 9:32  
CYCLE COMPLETE 24:32  
-----

Digital Signature #  
8EF135D28841E85F  
-----