

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2176

Name: RITTER 1  
Sterilizer ID: V2178198  
07/26/2022 09:25 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 183.7 F 0.0  
2:00 187.7 F 3.2  
4:00 189.2 F 7.9  
6:00 207.1 F 13.0  
8:00 220.6 F 7.6  
10:00 234.5 F 13.0  
12:00 235.9 F 10.0  
14:00 236.5 F 8.4  
16:00 247.7 F 14.5  
18:00 260.2 F 21.2  
20:00 271.1 F 28.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.6  
1:00 272.4 F 29.4  
2:00 272.3 F 29.0  
3:00 271.3 F 28.5  
4:00 272.1 F 29.3  
  
Min 271.2 F 28.4  
Max 272.7 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 213.1 F 0.0  
5:00 205.4 F 0.0  
10:00 213.6 F 0.0  
15:00 215.5 F 0.0  
20:00 217.9 F 0.0  
25:00 219.8 F 0.0  
30:00 217.9 F 0.0  
Drying Complete

FILLING: 1:51  
HEATING: 20:46  
STERILIZING: 4:00  
VENTING: 3:15  
DRYING: 30:00  
TOTAL CYCLE: 00:59:51  
CYCLE COMPLETE

07/26/2022 10:24 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_