
Midmark M11 - v1.0.5
Total Cycles: 1634

Name: RITTER2
Sterilizer ID: V2313228
10/21/2022 07:16 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 77.6 F 0.0
2:00 82.8 F 0.0
4:00 122.8 F 0.0
6:00 133.2 F 0.1
8:00 159.9 F 0.1
10:00 185.7 F 1.1
12:00 201.7 F 4.7
14:00 212.9 F 8.3
16:00 226.5 F 12.1
18:00 229.0 F 7.6
20:00 240.1 F 12.6
22:00 238.8 F 10.5
24:00 237.1 F 9.0
26:00 250.6 F 15.9
28:00 262.2 F 22.6
30:00 272.8 F 29.3

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.7
1:00 272.3 F 29.1
2:00 271.8 F 29.3
3:00 271.9 F 28.8
4:00 272.3 F 29.5

Min 271.3 F 28.6
Max 272.7 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.3 F 0.0
5:00 210.4 F 0.0
10:00 221.2 F 0.0
15:00 220.2 F 0.0
20:00 220.7 F 0.0
25:00 223.5 F 0.0
30:00 222.4 F 0.0
Drying Complete

FILLING: 0:49
HEATING: 30:19
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 01:08:18
CYCLE COMPLETE

10/21/2022 08:24 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____