
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
1.0uS / 0.6ppm
CYCLE NUMBER 001535
13:02 10/27/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:54
PRESSURIZING 5:54
STERILIZING 7:33
136.0 C 220kPa 7:33
136.4 C 223kPa 8:03
136.4 C 222kPa 8:33
136.3 C 221kPa 9:03
136.3 C 222kPa 9:33
136.3 C 223kPa 10:03
136.4 C 223kPa 10:33
136.3 C 221kPa 11:03
VENTING 11:03
AIR DRYING 12:07
CYCLE COMPLETE 42:07

Digital Signature #
3939A308E1B65813
