
Midmark M11 - v1.0.5
Total Cycles: 1800

Name: RITTER2
Sterilizer ID: V2313228
12/05/2022 10:49 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 155.1 F 0.0
2:00 154.9 F 1.9
4:00 161.4 F 4.9
6:00 183.1 F 9.3
8:00 210.9 F 6.8
10:00 223.3 F 8.7
12:00 234.3 F 10.9
14:00 239.0 F 11.7
16:00 238.5 F 10.3
18:00 251.8 F 17.0
20:00 263.5 F 23.9
22:00 271.5 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.3
1:00 272.3 F 29.7
2:00 271.6 F 29.6
3:00 272.1 F 29.6
4:00 271.5 F 29.5

Min 271.3 F 29.1
Max 272.4 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.6 F 0.0
5:00 233.0 F 0.0
10:00 233.2 F 0.0
15:00 229.9 F 0.0
20:00 228.9 F 0.0
25:00 229.2 F 0.0
30:00 232.6 F 0.0
Drying Complete

FILLING: 0:59
HEATING: 22:04
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 01:00:10
CYCLE COMPLETE

12/05/2022 11:49 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____