
Midmark M11 - v1.0.5
Total Cycles: 2122

Name: RITTER2
Sterilizer ID: V2313228
03/06/2023 01:07 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 176.3 F 0.0
2:00 189.3 F 3.8
4:00 212.0 F 8.1
6:00 228.7 F 10.9
8:00 231.8 F 8.8
10:00 237.2 F 10.3
12:00 242.5 F 12.3
14:00 241.4 F 11.4
16:00 255.1 F 18.3
18:00 267.0 F 25.6

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.9
1:00 272.1 F 29.2
2:00 272.3 F 29.6
3:00 271.8 F 28.9
4:00 272.3 F 29.6

Min 271.5 F 28.8
Max 272.5 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.3 F 0.0
5:00 223.4 F 0.0
10:00 242.1 F 0.0
15:00 233.2 F 0.0
20:00 226.0 F 0.0
25:00 225.6 F 0.0
30:00 227.1 F 0.0

Drying Complete

FILLING: 1:17
HEATING: 19:25
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:57:55
CYCLE COMPLETE

03/06/2023 02:04 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____