
Midmark M11 - v1.0.0
Total Cycles: 4721

Name: RITTER 1
Sterilizer ID: V2178198
07/12/2024 07:58 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.6 F 0.0
2:00 100.1 F 0.0
4:00 90.0 F 1.0
6:00 119.9 F 3.1
8:00 143.9 F 5.4
10:00 170.7 F 8.3
12:00 191.4 F 11.6
14:00 203.1 F 3.7
16:00 216.9 F 6.9
18:00 225.3 F 10.8
20:00 226.9 F 6.7
22:00 237.4 F 10.4
24:00 235.4 F 8.1
26:00 246.3 F 13.4
28:00 257.3 F 19.3
30:00 267.2 F 25.5

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.3
1:00 272.4 F 29.0
2:00 271.4 F 28.3
3:00 272.4 F 29.0
4:00 271.6 F 28.2

Min 271.4 F 28.2
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 189.0 F 0.0
5:00 218.6 F 0.0
10:00 214.8 F 0.0
15:00 232.1 F 0.0
20:00 212.6 F 0.0
25:00 199.7 F 0.0
30:00 195.1 F 0.0
Drying Complete

FILLING: 0:36
HEATING: 31:31
STERILIZING: 4:00
VENTING: 3:29
DRYING: 30:00
TOTAL CYCLE: 01:09:35
CYCLE COMPLETE

07/12/2024 09:07 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____