
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000070
12:21 07/02/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:14
PRESSURIZING 3:20
STERILIZING 4:38
136.4 C 220kPa 4:38
136.2 C 221kPa 5:08
136.4 C 222kPa 5:38
136.4 C 224kPa 6:08
136.4 C 225kPa 6:38
136.2 C 223kPa 7:08
136.4 C 223kPa 7:38
136.4 C 223kPa 8:08
136.4 C 223kPa 8:08
VENTING 8:09
AIR DRYING 8:59
CYCLE COMPLETE 35:21

Digital Signature #
B8C6B378576D064D
