

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5055

Name: RITTER 1  
Sterilizer ID: V2178198  
11/05/2024 09:28 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 70.4 F 0.0  
2:00 74.1 F 0.0  
4:00 105.6 F 0.0  
6:00 120.0 F 0.1  
8:00 148.4 F 0.2  
10:00 176.7 F 0.5  
12:00 195.9 F 3.0  
14:00 209.0 F 6.0  
16:00 218.9 F 9.3  
18:00 229.2 F 12.9  
20:00 228.9 F 7.6  
22:00 240.0 F 12.3  
24:00 237.2 F 9.3  
26:00 239.2 F 9.5  
28:00 247.1 F 13.5  
30:00 258.4 F 19.6  
32:00 268.8 F 26.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.0  
1:00 272.7 F 28.7  
2:00 271.6 F 27.9  
3:00 272.7 F 28.9  
4:00 272.1 F 28.3  
  
Min 271.5 F 27.9  
Max 272.8 F 28.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 185.9 F 0.0  
5:00 200.4 F 0.0  
10:00 198.0 F 0.0  
15:00 208.3 F 0.0  
20:00 216.7 F 0.0  
25:00 225.6 F 0.0  
30:00 231.9 F 0.0  
Drying Complete

FILLING: 0:42  
HEATING: 33:14  
STERILIZING: 4:00  
VENTING: 3:25  
DRYING: 30:00  
TOTAL CYCLE: 01:11:20  
CYCLE COMPLETE

11/05/2024 10:39 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_