
Midmark M11 - v1.0.5
Total Cycles: 2611

Name: RITTER2
Sterilizer ID: V2313228
07/05/2023 02:04 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 195.9 F 0.0
2:00 191.1 F 3.5
4:00 194.0 F 8.2
6:00 213.0 F 11.4
8:00 225.7 F 9.0
10:00 236.4 F 10.3
12:00 241.4 F 12.4
14:00 240.9 F 11.3
16:00 254.0 F 18.0
18:00 265.5 F 24.9

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.0
1:00 272.6 F 29.8
2:00 271.9 F 29.1
3:00 271.5 F 29.3
4:00 272.7 F 29.8

Min 271.2 F 28.8
Max 272.7 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.2 F 0.0
5:00 223.0 F 0.0
10:00 227.1 F 0.0
15:00 230.1 F 0.0
20:00 229.4 F 0.0
25:00 229.5 F 0.0
30:00 229.7 F 0.0
Drying Complete

FILLING: 1:59
HEATING: 19:48
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:58:58
CYCLE COMPLETE

07/05/2023 03:02 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____