
Midmark M11 - v1.0.5
Total Cycles: 3379

Name: RITTER2
Sterilizer ID: V2313228
02/13/2024 08:55 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 68.2 F 0.0
2:00 78.2 F 0.0
4:00 110.6 F 0.0
6:00 104.4 F 1.1
8:00 139.9 F 3.8
10:00 171.8 F 6.9
12:00 188.4 F 10.8
14:00 204.8 F 3.9
16:00 219.4 F 6.4
18:00 232.1 F 10.2
20:00 232.0 F 8.8
22:00 240.1 F 11.6
24:00 238.1 F 9.8
26:00 251.1 F 16.2
28:00 262.4 F 22.7
30:00 272.7 F 29.5

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.8
1:00 272.4 F 29.2
2:00 272.3 F 29.5
3:00 272.0 F 28.8
4:00 272.7 F 29.6

Min 271.5 F 28.7
Max 272.9 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 177.2 F 0.0
5:00 194.7 F 0.0
10:00 189.7 F 0.0
15:00 187.1 F 0.0
20:00 184.9 F 0.0
25:00 182.6 F 0.0
30:00 180.0 F 0.0
Drying Complete

FILLING: 0:49
HEATING: 30:22
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 01:08:21
CYCLE COMPLETE

02/13/2024 10:03 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____