

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3743

Name: RITTER2  
Sterilizer ID: V2313228  
06/28/2024 08:13 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 70.6 F 0.0  
2:00 85.5 F 0.0  
4:00 79.4 F 0.2  
6:00 118.4 F 1.1  
8:00 150.8 F 4.2  
10:00 172.6 F 7.9  
12:00 194.0 F 12.2  
14:00 214.7 F 4.4  
16:00 226.4 F 8.0  
18:00 237.2 F 12.2  
20:00 234.0 F 8.6  
22:00 242.1 F 11.5  
24:00 244.8 F 12.8  
26:00 257.1 F 19.4  
28:00 268.1 F 26.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 28.7  
1:00 272.3 F 28.9  
2:00 272.8 F 29.4  
3:00 271.8 F 28.9  
4:00 272.4 F 28.9

Min 271.7 F 28.5  
Max 272.9 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 192.4 F 0.0  
5:00 212.4 F 0.0  
10:00 186.0 F 0.0  
15:00 186.2 F 0.0  
20:00 185.2 F 0.0  
25:00 185.2 F 0.0  
30:00 183.9 F 0.0  
Drying Complete

FILLING: 0:51  
HEATING: 29:14  
STERILIZING: 4:00  
VENTING: 3:09  
DRYING: 30:00  
TOTAL CYCLE: 01:07:15  
CYCLE COMPLETE

06/28/2024 09:20 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_