
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 002470
17:12 28/01/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:25
PRESSURIZING 2:30
STERILIZING 3:34
136.0 C 222kPa 3:34
136.2 C 224kPa 4:04
136.1 C 223kPa 4:34
136.4 C 224kPa 5:04
136.4 C 227kPa 5:34
136.4 C 227kPa 6:04
136.4 C 227kPa 6:34
136.4 C 226kPa 7:04
136.4 C 226kPa 7:04
VENTING 7:04
AIR DRYING 8:03
CASSETTE REMOVED
CYCLE COMPLETE 24:08

Digital Signature #
31ACF1EC5F4032C0
