
Midmark M11 - v1.0.0
Total Cycles: 3060

Name: RITTER 1
Sterilizer ID: V2178198
03/30/2023 01:04 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 203.2 F 0.0
2:00 196.1 F 4.8
4:00 205.9 F 10.0
6:00 222.2 F 6.0
8:00 233.5 F 10.6
10:00 234.6 F 7.6
12:00 241.8 F 11.2
14:00 246.5 F 13.2
16:00 259.2 F 20.2
18:00 270.6 F 27.5

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.1
1:00 272.4 F 28.8
2:00 271.7 F 28.2
3:00 271.5 F 28.5
4:00 272.4 F 28.8

Min 271.1 F 28.0
Max 272.5 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.5 F 0.0
5:00 223.3 F 0.0
10:00 234.2 F 0.0
15:00 230.0 F 0.0
20:00 229.9 F 0.0
25:00 227.5 F 0.0
30:00 226.3 F 0.0

Drying Complete

FILLING: 1:53
HEATING: 18:49
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 00:57:59
CYCLE COMPLETE

03/30/2023 02:01 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____