

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5368

Name: RITTER 1  
Sterilizer ID: V2178198  
02/26/2025 12:27 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 173.1 F 0.0  
2:00 170.4 F 3.0  
4:00 178.0 F 7.0  
6:00 195.7 F 11.3  
8:00 213.7 F 3.5  
10:00 223.5 F 7.8  
12:00 234.6 F 12.1  
14:00 234.5 F 8.5  
16:00 244.6 F 12.4  
18:00 244.1 F 12.2  
20:00 255.4 F 18.2  
22:00 265.8 F 24.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.3  
1:00 272.4 F 29.1  
2:00 272.1 F 28.8  
3:00 271.5 F 28.3  
4:00 271.4 F 28.6

Min 271.2 F 28.2  
Max 272.6 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 203.8 F 0.0  
5:00 207.0 F 0.0  
10:00 220.1 F 0.0  
15:00 230.4 F 0.0  
20:00 219.8 F 0.0  
25:00 223.0 F 0.0  
30:00 227.9 F 0.0  
Drying Complete

FILLING: 1:33  
HEATING: 23:53  
STERILIZING: 4:00  
VENTING: 3:58  
DRYING: 30:00  
TOTAL CYCLE: 01:03:24  
CYCLE COMPLETE

02/26/2025 01:30 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_