

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1844

Name: RITTER2  
Sterilizer ID: V2313228  
12/15/2022 08:58 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 173.5 F 0.0  
2:00 171.3 F 2.1  
4:00 178.9 F 6.2  
6:00 202.6 F 12.1  
8:00 221.1 F 7.0  
10:00 234.7 F 12.6  
12:00 237.1 F 10.3  
14:00 236.7 F 9.0  
16:00 250.7 F 16.2  
18:00 263.1 F 23.5  
20:00 271.8 F 29.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 29.2  
1:00 272.6 F 29.5  
2:00 272.2 F 29.7  
3:00 272.4 F 29.4  
4:00 272.2 F 29.7

Min 271.8 F 29.0  
Max 272.9 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 204.9 F 0.0  
5:00 229.2 F 0.0  
10:00 229.2 F 0.0  
15:00 224.9 F 0.0  
20:00 216.9 F 0.0  
25:00 215.7 F 0.0  
30:00 212.3 F 0.0  
Drying Complete

FILLING: 1:58  
HEATING: 20:02  
STERILIZING: 4:00  
VENTING: 3:10  
DRYING: 30:00  
TOTAL CYCLE: 00:59:11  
CYCLE COMPLETE

12/15/2022 09:57 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_